

PLASMA BIONICS

MATERIAL COMPATIBILITY WITH AIR PLASMA STERILIZATION™

COMPATIBILITY WITH COMMON PACKAGING

The Air Plasma Sterilization™ process is compatible with commonly used sterile barrier packaging. This includes SMS wraps, Tyvek® pouches, and Tyvek®-Mylar® pouches. The process is not compatible with non-porous cellulosic materials such as paper, cardboard, and textiles (e.g. cloth and linen). Some pouches designed for steam and ethylene oxide may become brittle when used with Air Plasma Sterilization™.

COMPATIBILITY WITH MEDICAL DEVICE MATERIALS

Ozone (O3), nitrogen dioxide (NO2), and other reactive oxygen and nitrogen species are powerful oxidants that can be used to sterilize many materials. However, the high oxidation potentials of these sterilants may result in degradation of some materials. The Air Plasma Sterilization™ process has been tested with different materials commonly used in medical practices including metals, polymers, glass, and plastics. Some mild discoloration or yellowing may occur with some dyed elastomers and plastics.

The following lists of compatible and incompatible materials are not exhaustive. For inquiries about compatibility of specific materials, please contact info@plasmabionics.com.

COMPATIBLE MATERIALS

<u>METALS</u>	<u>PLASTICS</u>		<u>ELASTOMERS</u>	<u>OTHER</u>
Stainless steel (300 series, medical grade) Titanium Anodized aluminum Gold plating	Polytetrafluoroethylene (PTFE, Teflon) Polyvinylidene difluoride (PVDF, Kynar) Acrylonitrile butadiene styrene (ABS) Polybutylene terephthalate (PBT) Polyethylene (PE, HDPE, UHMWPE) Polyether ether ketone (PEEK)	Polyvinylchloride (PVC) Polypropylene (PP) Polycarbonate (PC) Polyurethane (PU) Polysulfone (PSU)	Silicone* Fluorosilicone Viton (FKM) Santoprene Tygon Ethylene propylene diene monomer (EPDM)	Glass Ceramic

INCOMPATIBLE MATERIALS

<u>METALS</u>			<u>PLASTICS</u>	<u>ELASTOMERS</u>	<u>OTHER</u>
Mild steel Galvanized steel Brass	Zinc Tin	Nickel Bronze Copper	Nylon Delrin (Acetal) Polyamide	Natural rubber Buna-N (Nitrile)	Cellulosic material (paper, cardboard, textiles)

*There are numerous blends, grades, and varieties of silicone. It is advised to test specific blends prior to routine use.